



**FINISHED PACK SUMMARY**

BATCH NUMBER 9001

FPS M47xCE  
Issue Date: 04/2009

TITLE: MICROSCREEN STREP EXTRACTION ENZYME - CE

**MICROGEN BIOPRODUCTS**

**CERTIFICATE OF CONFORMITY**

Name of Product: MICROSCREEN STREP EXTRACTION ENZYME  
Product Code: M47xCE  
Components: One of each: M47xCE

1. This component complies with current Microgen Bioproducts approved specifications.
2. Component performance complies with current release specifications.

**KEY DATA SUMMARY**

Component	Part Number	Lot	Expiry	Lot, expiry correct SOP 340	Component PMP reviewed SOP 337
Ext Enz	M47xCE	9001	2017-05	<i>Pass</i>	<i>Pass</i>
Assembly	M47xCE	9001	2017-05	<i>Pass</i>	<i>Pass</i>

Overall Expiry Date: 2017-05

Component	Appearance	Fill volume	Sodium Azide	Function
Extraction enzyme - M47xCE	Freeze-dried reagent	Reconstituted to 10ml	Negative - Contains Thiomersal	Lysis >60%

**COMPLIES**

SIGNATURE: *P. Hayden*

DATE: 22-DEC-15