



MICROGEN BIOPRODUCTS

CERTIFICATE OF CONFORMITY

Name of Product:	Microgen GN-ID Panel A Kit 60 test
Product Code:	MID-64
Active Components:	10 x MID 64A each containing 6 tests

1. All components in this kit comply with current Microgen Bioproducts approved specifications.

2. Performance of strips with QC test panel complies with current release specifications:

K oxytoca MBCC 46
 E cloacae MBCC 51
 S marcescens MBCC 59
 E coli MBCC 72
 P mirabilis MBCC 78
 P stuartii MBCC 168
 S typhimurium MBCC 215
 P aeruginosa MBCC 169

Give expected pattern of substrate reactions

3. Product satisfies final Quality Control examination according to FPS MID-64

Kit Lot No.	38951
Expiry	2017-06

SIGNATURE:	<i>P. Walker</i>
DATE:	17-DEC-15