



MICROGEN BIOPRODUCTS
CERTIFICATE OF CONFORMITY

Name of Product: Microgen GN-ID Panel B Kit 24 test
 Product Code: MID-65
 Active
 Components: 4 x MID 65A pouches each containing 6 tests

1. All components in this kit comply with current Microgen Bioproducts approved specifications.
2. Performance of strips with QC test panel complies with current release specifications:

K oxytoa MBCC 46
 E cloacae MBCC 51
 S marcescens MBCC 59
 E coli MBCC 72
 P mirabilis MBCC 78
 P stuartii MBCC 168
 S typhimurium MBCC 215
 P aeruginosa MBCC 169

Give expected pattern of substrate reactions

3. Product satisfies final Quality Control examination according to FPS MID-65

Kit Lot No.

08451

Expiry

2017-05

SIGNATURE:

[Handwritten Signature]

DATE:

16-NOV-15