



**MICROGEN BIOPRODUCTS**  
**CERTIFICATE OF CONFORMITY**

Name of Product: Microgen Streptococcus ID 20 test  
Product Code: MID-62  
Components: 20 x MID 62C; 20 x MID62B 1 x MID62D

1. All components in this kit comply with current Microgen Bioproducts approved specifications.

2. Performance of strips with QC test panel complies with current release specifications:

*Streptococcus mutans* MBCC 1067  
*Streptococcus uberis* MBCC 1072  
*Gemella haemolysans* MBCC 1085

*Enterococcus avium* MBCC 1136  
*Enterococcus faecalis* MBCC 1137

Give expected pattern of substrate reactions.

3. Product satisfies final Quality Control examination according to FPS MID-62

Component	Inventory Code	Lot Number	Expiry	Lot-expiry correct SOP340	Component PMP reviewed SOP337
Suspending Broth	MID62B	2701	2019-04	PASS	PASS
Test Strip	MID62C	2601	2018-09	PASS	PASS
Hippurate	MID62D	2701	2018-09	PASS	PASS
Assembly	MID62	02771	2018-09	PASS	PASS
Overall Expiry Date: 2018-09					

SIGNATURE: *RAV*

DATE: 22 / MAY / 2017