



## MICROGEN BIOPRODUCTS

### CERTIFICATE OF CONFORMITY

Name of Product: Microgen Listeria ID 20 test  
 Product Code: MID-67  
 Components: 20 x MID 67C; 20 x MID67B; 1 x MID67A

1. All components in this kit comply with current Microgen Bioproducts approved specifications.  
 2. Performance of strips with QC test panel complies with current release specifications:

L monocytogenes MBCC305	L. seeligeri MBCC310
L grayii MBCC307	L. welshimeri MBCC308
L. innocua MBCC306	L. ivanovii MBCC325

Give expected pattern of substrate reactions.

3. Product satisfies final Quality Control examination according to FPS MID-67

Component	Inventory Code	Lot Number	Expiry	Lot-expiry correct SOP340	Component PMP reviewed SOP337
Haemolysin Reagent	MID67A	12901C	2018-02	PASS	PASS
Suspending Broth	MID67B	35701	2018-09	PASS	PASS
Test Strip	MID67C	35701	2018-10	PASS	PASS
Assembly	MID67	<del>35701</del>	2018-02	PASS	PASS
<b>Overall Expiry Date:</b> 2018-02					

SIGNATURE:	
DATE: 03 / 09 / 2017	