

# MICROGEN BIOPRODUCTS LIMITED

1 Admiralty Way  
Camberley  
Surrey GU15 3DT  
United Kingdom

Telephone: + 44 (0)1276 600 081  
Fax: + 44 (0)1276 600 151



## CERTIFICATE OF ANALYSIS

**Description:** Path-Chek Listeria Media PC080      **Lot:** 53809      **Expiry:** 2018-09

**Storage:** 2 — 8 degrees C      **Do not freeze**

**Shipping:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures

---

TEST PARAMETERS:	SPECIFICATION:	RESULT:
Volume :	3.0 +/- 0.2mL	Pass
pH:	7.2 +/- 0.2	Pass
Sensitivity against <i>Listeria monocytogenes</i> :	<10	Pass
Reaction with <i>E coli</i>	negative	Pass
Sterility:	No growth	Pass

Signed:



Quality Assurance

Date:

13/09/2017