

**MICROGEN**  
BIOPRODUCTS

***MICROGEN BIOPRODUCTS***

***CERTIFICATE OF CONFORMITY***

Name of Product:	Microgen MID-644 GN A+B ID System 30 test kit
Product Code:	MID-644CE
Active Components:	5 x MID 64A each containing 6 tests 5 x MID 65A each containing 6 tests

1. All components in this kit comply with current Microgen Bioproducts approved specifications.

2. Performance of strips with QC test panel complies with current release specifications:

K oxytoca MBCC 46  
E cloacae MBCC 51  
S marcescens MBCC 59  
E coli MBCC 72  
P mirabilis MBCC 78  
P stuartii MBCC 168  
S typhimurium MBCC 215  
P aeruginosa MBCC 169

Give expected pattern of substrate reactions

3. Product satisfies final Quality Control examination according to FPS MID-644

MID-64a Lot No. 47901	Expiry 2020-04
MID-65a Lot No. 10401	Expiry 2020-03

MID-644ce Lot No. 10181	Expiry 2020-03
-------------------------	----------------

SIGNATURE: <i>Sajana</i>
DATE: 10 Oct '18