

MICROGEN
BIOPRODUCTS

MICROGEN BIOPRODUCTS

CERTIFICATE OF CONFORMITY

Name of Product:	Microgen MID-644 GN A+B ID System 30 test kit
Product Code:	MID-644CE
Active Components:	5 x MID 64A each containing 6 tests 5 x MID 65A each containing 6 tests

1. All components in this kit comply with current Microgen Bioproducts approved specifications.

2. Performance of strips with QC test panel complies with current release specifications:

K oxytoca MBCC 46
E cloacae MBCC 51
S marcescens MBCC 59
E coli MBCC 72
P mirabilis MBCC 78
P stuartii MBCC 168
S typhimurium MBCC 215
P aeruginosa MBCC 169

Give expected pattern of substrate reactions

3. Product satisfies final Quality Control examination according to FPS MID-644

MID-64a Lot No. 48101	Expiry 2020-05
MID-65a Lot No. 10501	Expiry 2020-05

MID-644ce Lot No. 10281	Expiry 2020-05
-------------------------	----------------

SIGNATURE: Sajana
DATE: 29 Nov 18