

MICROGEN BIOPRODUCTS LIMITED

Unit 1, Watchmoor Point

Camberley

Surrey GU15 3AD

United Kingdom

Telephone: + 44 (0)1276 600 081

Fax: + 44 (0)1276 600 151

MICROGEN
BIOPRODUCTS

CERTIFICATE OF CONFORMITY

Description: Path-Chek Listeria Media **Lot No:** 58821 **Expiry:** 2020-07
Storage: 2 — 8 ° C **Do not freeze**
Shipping: This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures

TEST PARAMETERS:	SPECIFICATION:	RESULT:
Volume	3.0 +/- 0.2mL	Pass
pH	7.2 +/- 0.2	Pass
Sensitivity against <i>Listeria monocytogenes</i>	< 10	Pass
Reaction with <i>E coli</i>	Negative	Pass
Sterility	No Growth	Pass

Signed:

Sajana

Quality Assurance

Date:

09 Aug 19