

# MICROGEN BIOPRODUCTS LIMITED

Unit 1, Watchmoor Point  
Camberley  
Surrey GU15 3AD  
United Kingdom

Telephone: + 44 (0)1276 600 081  
Fax: + 44 (0)1276 600 151

MICROGEN  
BIOPRODUCTS

## CERTIFICATE OF CONFORMITY

**Description:** Path-Chek Listeria Media      **Lot No:** 59948      **Expiry:** 2020-12  
**Storage:** 2 — 8 ° C      **Do not freeze**  
**Shipping:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures

---

TEST PARAMETERS:	SPECIFICATION:	RESULT:
Volume	3.0 +/- 0.2mL	Pass
pH	7.2 +/- 0.2	Pass
Sensitivity against <i>Listeria monocytogenes</i>	< 10	Pass
Reaction with <i>E coli</i>	Negative	Pass
Sterility	No Growth	Pass

Signed:  Quality Assurance

Date: 17/Jan/20