MICROGEN BIOPRODUCTS LIMITED

Unit 1, Watchmoor Point
Camberley
Surrey GU15 3AD
United Kingdom

Telephone: + 44 (0)1276 600 081
Fax: + 44 (0)1276 600 151

MICROGEN
BIOPRODUCTS

CERTIFICATE OF CONFORMITY

Description: **Pathchek Hygiene Protein** **Product Code PC006** **Lot No. PW5L161I**

Expiry Date: 2021-12

Storage: **4 — 25 degrees C** **Do not freeze**

Shipping: **This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures**

TEST PARAMETERS:	SPECIFICATION:	RESULT:
Sensitivity	Detects 0.02mg Bovine Serum Albumin	PASS
Instructions For Use	Current Version present	PASS
Packaging	Silver Foil Pouch bearing product label With Lot Number and Expiry Date Each pouch containing 5 x 10 double ended swabs	PASS

Signed: Dawn Hi

Quality Assurance

Date: 31/Jan/20