

MICROGEN

BIOPRODUCTS

CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Flu A & B Rapid Test Cassette (Swab/ Nasal Aspirate)	FLU20030052	M591CE
EXPIRY DATE:		2022-03

TEST PARAMETERS	SPECIFICATION	RESULT
Appearance	Good	Pass
Positive Samples	100% Positive	Pass
Negative Samples	100% Negative	Pass

Storage: Store at room temperature or refrigerated between 2 – 30°C.
Do not freeze.


We hereby certify that this material/product has been released as meeting our acceptance criteria

Signed:



Date: 21 April 2021

Quality Assurance

 Microgen Bioproducts Ltd.
Unit 1 Watchmoor Point
Camberley
Surrey GU15 3AD
UK
Part of the Novacyt Group

TEL: +44 (0) 1276 600081
FAX: +44 (0) 1276 600151

enquiries.microgen@novacyt.com

2021-01

MB.PF.FLUAB.COA V1