

**MICROGEN BIOPRODUCTS**

**CERTIFICATE OF CONFORMITY**

Name of Product: MICROGEN Staph  
 Product Code: M43(CE)  
 Components: One of each:- M43a, M43b

1. All components in this kit comply with current Microgen Bioproducts approved specifications.
2. Performance of kit complies with current release specifications.
3. Performance of test latex with QC panel gives expected pattern of reactions and complies with current release specifications.
 

S. aureus (MRSA) MBCC15, positive Staph aureus MBCC193, positive Staph aureus MBCC194, positive Staph aureus MBCC195, positive Staph aureus MBCC197, positive S. aureus (MRSA) MBCC568, positive Staph aureus MBCC693, positive	Staph epid. MBCC198, negative Staph epid. MBCC199, negative Saline, negative
---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	------------------------------------------------------------------------------------
4. All components containing human product have been tested for HIV 1/2 Ab, HCV and HbsAg as individual donors and were found to be negative.

**KEY DATA SUMMARY**

Component	Part Number	Lot	D.O.M	Expiry	Lot, expiry correct SOP 340	Component PMP reviewed SOP 337
Latex reagent	M43a	17001	2021-04	2022-12	PASS	PASS
Positive Control	M43b	20201	2021-04	2022-10	PASS	PASS
Assembly	M43	14411	2021-04	2022-10	PASS	PASS

Overall Expiry Date:	2022-10
----------------------	---------

SIGNATURE: <i>Diana W</i>
DATE: 12/ May 21