

CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Flu A & B Rapid Test Cassette (Swab/ Nasal Aspirate)	FLU19120020	M591CE
EXPIRY DATE:		2021-12

TEST PARAMETERS	SPECIFICATION	RESULT
Appearance	Good	Pass
Positive Samples	100% Positive	Pass
Negative Samples	100% Negative	Pass

Storage: Store at room temperature or refrigerated between 2 – 30°C.
Do not freeze.

We hereby certify that this material/product has been released as meeting our acceptance criteria

Signed:



Date: 07 JULY 2021

Quality Assurance