

CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Microgen Bioproducts Listeria ID 20 Test Kit	312041	MID67
EXPIRY DATE:		2022-08

Component	Lot Number	Expiry
Haemolysin Reagent	17001B	2022-08
Suspending Broth	312031	2023-03
Test Strip	295051	2023-04

Species	
L. monocytogenes MBCC305	Technical performance complies with current release specifications
L. seeligeri MBCC310	
L. grayi MBCC307	
L. welshimeri MBCC308	
L. innocua MBCC306	
L. ivanovii MBCC325	

Storage: Store refrigerated at 2 – 8°C.
Do not freeze

Shipping: This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures.

We hereby certify that this material/product has been released as meeting our acceptance criteria

Signed:



Date:

19/NOV/21

Quality Assurance