

## CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Microgen Bioproducts Listeria ID 20 Test Kit	333071	MID67
<b>EXPIRY DATE:</b>		2022-08

Component	Lot Number	Expiry
Haemolysin Reagent	17001E	2022-08
Suspending Broth	333061	2023-01
Test Strip	330011	2023-05

Species	
L monocytogenes MBCC305	Technical performance complies with current release specifications
L. seeligeri MBCC310	
L grayi MBCC307	
L. welshimeri MBCC308	
L. innocua MBCC306	
L. ivanovii MBCC325	

**Storage:** Store refrigerated at 2 – 8°C.  
Do not freeze

**Shipping:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures.

---

**We hereby certify that this material/product has been released as meeting our acceptance criteria**

---

**Signed:** 

**Date:** 09/DEC/21

Quality Assurance