

CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Microgen Bioproducts Listeria ID 20 Test Kit	025022	MID67
EXPIRY DATE:		2022-10

Component	Lot Number	Expiry
Haemolysin Reagent	287011	2022-10
Suspending Broth	025012	2023-04
Test Strip	011052	2023-07

Species	
L monocytogenes MBCC305	Technical performance complies with current release specifications
L. seeligeri MBCC310	
L grayi MBCC307	
L. welshimeri MBCC308	
L. innocua MBCC306	
L. ivanovii MBCC325	

Storage: Store refrigerated at 2 – 8°C.
Do not freeze

Shipping: This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures.

We hereby certify that this material/product has been released as meeting our acceptance criteria

Signed:



Date:

04/FEB/2022

Quality Assurance