

# MICROGEN BIOPRODUCTS

## CERTIFICATE OF ANALYSIS

| DESCRIPTION                                  | LOT NO. | PRODUCT CODE |
|--|---------|--------------|
| Microgen Bioproducts Listeria ID 20 Test Kit | 53011   | MID67        |
| <b>EXPIRY DATE:</b>                          |         | 2022-06      |

| Component          | Lot Number | Expiry  |
|--------------------|------------|---------|
| Haemolysin Reagent | 16801D     | 2022-06 |
| Suspending Broth   | 53001      | 2023-01 |
| Test Strip         | 53001      | 2023-01 |

| Species                 | Technical performance complies with current release specifications |
|-------------------------|--|
| L monocytogenes MBCC305 |  |
| L. seeligeri MBCC310    |  |
| L grayi MBCC307         |  |
| L. welshimeri MBCC308   |  |
| L. innocua MBCC306      |  |
| L. ivanovii MBCC325     |  |

**Storage:** Store refrigerated at 2 – 8°C.  
Do not freeze

**Shipping:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures.

---

**We hereby certify that this material/product has been released as meeting our acceptance criteria**

---

**Signed:**

*puetania*

**Date:**

*17/DEC/21*

Quality Assurance