

## CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Microgen Bioproducts Listeria ID 20 Test Kit	054022	MID67
<b>EXPIRY DATE:</b>		2022-10

Component	Lot Number	Expiry
Haemolysin Reagent	287011	2022-10
Suspending Broth	054012	2023-03
Test Strip	049072	2023-08

Species	
L monocytogenes MBCC305	Technical performance complies with current release specifications
L. seeligeri MBCC310	
L grayi MBCC307	
L. welshimeri MBCC308	
L. innocua MBCC306	
L. ivanovii MBCC325	

**Storage:** Store refrigerated at 2 – 8°C.  
Do not freeze

**Shipping:** This product has data supporting stability tolerance during fluctuations in ambient shipping temperatures.

---

**We hereby certify that this material/product has been released as meeting our acceptance criteria**

---

**Signed:** 

**Date:** 03/MAR/2022

Quality Assurance