

CERTIFICATE OF ANALYSIS

DESCRIPTION	LOT NO.	PRODUCT CODE
Rotavirus and Adenovirus Combo Rapid Test Cassette (Feces)	RAC22040022	M595CE
EXPIRY DATE:		2024-03

TEST PARAMETERS		SPECIFICATION	RESULT
Appearance		Good	Pass
ROT	Positive Samples	100% Positive	Pass
	Negative Samples	100% Negative	Pass
ADE	Positive Samples	100% Positive	Pass
	Negative Samples	100% Negative	Pass

Storage: Store at room temperature or refrigerated between 2 – 30°C.
Do not freeze.

We hereby certify that this material/product has been released as meeting our acceptance criteria

Signed: 

Date: 09/MAY/2022

Quality Assurance